

# **CORRIGENDUM**

*Dated:25-02-2015*

## **I. The dates are new as Follows:-**

- 1. Submission of Forms – 05-03-2015**
- 2. Opening of Forms – 08-03-2015**

## **II. Changes in the Specifications:-**

1. In the requirement the line “Detector Should be 128 rows or more” is deleted.
2. (4). (b) Anode heat Storage Capacity of atleast 7.5 MHu should be read as Anode Heat Storage Capacity of at least 7.0 MHu.
3. Number of Physically independent row of detectors must be 128 or more is changed to number of physically independent row of detectors must be 64 or more.

## **III. In case of Tie for the rent.**

### **The Criteria for Awarding the Tender will be as follows:-**

- (A). First Preference would be given to that bidder who offers more concession for BPL/DDY/ Doctors of M.Y. Hospital then the concession specified in the tender.
- (B). In case of Tie of A then the preference would be given to that bidder who offers lesser rate for various investigations than the Rate specified in the Tender.

Jt. Director Cum Superintendent  
M.Y. Hospital, Indore